

Certificate of Analysis

Product name: 100 mm Dish
Product type: Oosafe® Plasticware IVD
Product code: OOPW-HD10
Batch/LOT No.: 09004
Product expiry date: 2029-06

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-4670-0724
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	S8108-0724SPAR
Limulus Amoebocyte Lysate (LAL)	<0.03 EU/device	Passed	Do not contain endotoxin	E4124-0724SPAR

Quality Management system according to ISO 13485:2016


Sterilized by irradiation according to ISO 11137-1

Sterilization method: E-Beam




Sterility assurance level: 10⁻⁶

GOosafe with SparMED!

Manuela Jagieniak
 Quality Control Department
 SparMED ApS
 Issue date: 29.07.2024



Sparmed ApS

Ryttermarken 2, 3520 Farum, DK  +45 39 40 25 03  info@sparmed.dk  www.sparmed.dk
 Corp. VAT no.: DK 30 89 85 75