

# Certificate of Analysis

**Product name: Center Well Dish with 2 Compartments**

**Product type: Oosafe® Plasticware IVD**

**Product code: OOPW-CW03**

**Batch/LOT No.: 09006**

**Product expiry date: 2029-06**

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-4670-0724
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	S8118-0724SPAR
Limulus Amoebocyte Lysate (LAL)	<0.03 EU/device	Passed	Do not contain endotoxin	E4124-0724SPAR

Quality Management system according to ISO 13485:2016

**Sterilized by irradiation according to ISO 11137-1**

**Sterilization method: E-Beam**

**Sterility assurance level: 10<sup>-6</sup>**


**GOosafe with SparMED!**

Manuela Jagieniak

Quality Control Department

SparMED ApS

Issue date: 31.07.2024



**Sparmed ApS**

Ryttermarken 2, 3520 Farum, DK +45 39 40 25 03 info@sparmed.dk www.sparmed.dk  
Corp. VAT no.: DK 30 89 85 75