

Certificate of Analysis


Product name: 100 mm Dish
Product type: Oosafe® Plasticware IVD
Product code: OOPW-HD10
Batch/LOT No.: 09012
Product expiry date: 2029-08

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-4947-0924
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	S8215-0924SPAR
Limulus Amoebocyte Lysate (LAL)	<0.03 EU/device	Passed	Do not contain endotoxin	E4220-0924SPAR




Quality Management system according to ISO 13485:2016

Sterilized by irradiation according to ISO 11137-1
Sterilization method: E-Beam
Sterility assurance level: 10⁻⁶

GOosafe with SparMED!
 Manuela Jagieniak
 Quality Control Department
 SparMED ApS
 Issue date: 24.09.2024



Sparmed ApS

Ryttermarken 2, 3520 Farum, DK  +45 39 40 25 03  info@sparmed.dk  www.sparmed.dk
 Corp. VAT no.: DK 30 89 85 75