

Certificate of Analysis

Product name: Oosafe® Snap-Cap Tube, 14 ml

Product type: Oosafe® Plasticware IVD

Product code: OOPW-OT10

Batch/LOT No.: 09113

Product expiry date: 2030-06

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-1391-25-08
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	SPAR-1393-25-08
Limulus Amoebocyte Lysate (LAL)	<0.03 EU/device	Passed	Do not contain endotoxin	SPAR-1394-25-08

Quality Management system according to ISO 13485:2016

Sterilized by irradiation according to ISO 11137-1

Sterilization method: E-Beam

Sterility assurance level: 10⁻⁶


GOosafe with SparMED!

Manuela Jagieniak




Quality Control Department

SparMED ApS

Issue date: 27.08.2025



Sparmed ApS

Ryttermarken 2, 3520 Farum, DK  +45 39 40 25 03  info@sparmed.dk  www.sparmed.dk
Corp. VAT no.: DK 30 89 85 75