

# Certificate of Analysis

**Product name: 35mm Dish, Non-Treated**

**Product type: Oosafe® Plasticware**

**Product code: OOPW-TF06**

**Batch/LOT No.: 09206**

**Product expiry date: 2030-11**

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	TEST REPORT NO.
Mouse Embryo Assay (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	MEA.025.2548.2025
Human Sperm Assay (HSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	SSMA.025.2549.2025
Limulus Amoebocyte Lysate (LAL)	≤ 20 EU/device	Passed	Do not contain endotoxin	LAL.025.2550.2025

Quality Management system according to ISO 13485:2016

**Sterilized by irradiation according to ISO 11137-1**

**Sterilization method: E-Beam**

**Sterility assurance level: 10<sup>-6</sup>**

**GOosafe with SparMED!**

Manuela Jagieniak

Quality Control Department

SparMED ApS

Issue date: 30.03.2026



**Sparmed ApS**

Ryttermarken 2, 3520 Farum, DK +45 39 40 25 03 info@sparmed.dk www.sparmed.dk

Corp. VAT no.: DK 30 89 85 75