

Certificate of Analysis

Product name: Center Well Dish with 2 Compartments
Product type: Oosafe® Plasticware
Product code: OOPW-CW03
Batch/LOT No.: 09209
Product expiry date: 2030-11

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	TEST REPORT NO.
Mouse Embryo Assay (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	MEA.025.2545.2025
Human Sperm Assay (HSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	SSMA.025.2546.2025
Limulus Amoebocyte Lysate (LAL)	≤ 20 EU/device	Passed	Do not contain endotoxin	LAL.025.2547.2025

Quality Management system according to ISO 13485:2016

Sterilized by irradiation according to ISO 11137-1
Sterilization method: E-Beam
Sterility assurance level: 10⁻⁶

GOosafe with SparMED!
 Manuela Jagieniak
 Quality Control Department
 SparMED ApS
 Issue date: 30.03.2026

