



SUCCESS THROUGH INNOVATION

# CERTIFICATE OF ANALYSIS

<b>PRODUCT DESCRIPTION:</b>	EZ-Squeeze
<b>LOT NUMBER(S):</b>	201606, 201634
<b>EXPIRY DATE:</b>	2022-05
<b>INDICATIONS FOR USE:</b>	The EZ-Squeeze pipettes are for denudation, i.e. removing the cumulus from an oocyte prior to the assisted reproduction techniques of intracytoplasmic sperm injection (ICSI) and in vitro fertilization (IVF), and effecting dish to dish transfer of cells, embryos and oocytes.
<b>QUALITY ASSURANCE:</b>	<p>Sterilised by gamma irradiation using dose between 25 and 45kGy.</p> <p>Pipette tips non-embryotoxic proven by Mouse Embryo Assay Test. Pass criteria 2-cells at day 2 <math>\geq</math> 80%, expanded blastocysts at day 6 <math>\geq</math> 80%. <b>PASS.</b></p> <p>Pipette tips non-pyrogenic proven by Limulus Amoebocyte Lysate Test. Pass criteria <math>&lt;</math>20Eu/device. <b>PASS.</b></p>
<b>STORAGE CONDITIONS:</b>	Temperature 10-30°C Humidity 20-80% RH Keep away from direct sunlight (UV)

Quality Assurance:

Date: 21 July 2017

**R<sub>X</sub>**Only *Caution: US Federal law restricts this device to sale by or on the order of a physician.*

Research Instruments Ltd, Bickland Industrial Park, Falmouth, Cornwall, TR11 4TA, UK

Tel : +44 1326 372 753 Fax: +44 1326 378 783

[sales@research-instruments.com](mailto:sales@research-instruments.com)[www.research-instruments.com](http://www.research-instruments.com)