

Gynetics Medical Products CERTIFICATE OF QUALITY

ITEM name and code: #4164 Probet
Endometrial Curette

LOT NUMBER: 20170110

EXPIRY DATE: 2021-01-10

STERILIZATION DATE: 16-01-2017

STERILIZATION TECHNIQUE: EO

Results of sterility test with bio-indicators No growth of organisms

Gynetics guarantees the following results due to a validated EO sterilization process.

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATIONS	Result
Ethylene oxide residue testing	<250 ppm	Pass
Ethylene chlorhydrin residue testing	<250 ppm	Pass
Suction function testing	<-110 mbars	Pass



Els Rutten
Gynetics Medical Products N.V.

26-01-2017

Date