

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:

Soft Embryo Transfer Catheter Semtrac Set 4.2 #2000 Set 4.2

LOT number: 20170116**STERILIZATION date:** 21-01-2017**EXPIRY date:** 20210117**STERILIZATION technique:** ETO**STERILIZATION dose:** 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans
Gynetics Medical Products

SIGNATURE date:

08-02-2017