

# Gynetics Medical Products

## CERTIFICATE OF QUALITY

**ITEM name and code:**

Intra-Uterine Insemination Cannula Soft #4225 Smooze

**LOT number:** 20170117**STERILIZATION date:** 21-01-2017**EXPIRY date:** 20210117**STERILIZATION technique:** ETO**STERILIZATION dose:** 25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



K. Daemen  
Gynetics Medical Products

**SIGNATURE date:**

08-02-2017