

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:

Intra-Uterine Insemination Cannula #4220 IUI

LOT number: 20170215**STERILIZATION date:** 16-02-2017**EXPIRY date:** 15-02-2021**STERILIZATION technique:** Gamma irradiation**STERILIZATION dose:** 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



K. Daemen
Gynetics Medical Products

SIGNATURE date:

08-03-2017