

Gynetics Medical Products CERTIFICATE OF QUALITY

ITEM name and code: #4164 Probet
Endometrial Curette

LOT NUMBER: 20170329

EXPIRY DATE: 2021-03-29

STERILIZATION DATE: 1-04-2017

STERILIZATION TECHNIQUE: EO

Results of sterility test with bio-indicators No growth of organisms

Gynetics guarantees the following results due to a validated EO sterilization process.

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATIONS	Result
Ethylene oxide residue testing	<250 ppm	Pass
Ethylene chlorhydrin residue testing	<250 ppm	Pass
Suction function testing	<-110 mbars	Pass



Karen Daemen
Gynetics Medical Products N.V.



Date