

CERTIFICATE OF QUALITY

ITEM name and code	#4000 Tulip LS TIS - Embryo Transfer Catheter
LOT NUMBER	20170329
EXPIRY DATE	29 March 2021
STERILIZATION DATE	24 June 2017
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans
Gynetics Medical Products N.V.

Signature date:
6 September 2018