

## CERTIFICATE OF QUALITY

|                                |                                           |
|--------------------------------|-------------------------------------------|
| <b>ITEM name and code</b>      | #3145 Handling Pipette - Handling Pipette |
| <b>LOT NUMBER</b>              | 20170405                                  |
| <b>EXPIRY DATE</b>             | 5 April 2021                              |
| <b>STERILIZATION DATE</b>      | 8 April 2017                              |
| <b>STERILIZATION TECHNIQUE</b> | Gamma Irradiation                         |
| <b>STERILIZATION DOSE</b>      | 25 - 39 kGray                             |

## QUALITY CONTROL TESTING PROCEDURES

| TEST                  | SPECIFICATION  | CONCLUSION |
|-----------------------|----------------|------------|
| Endotoxin Assay (LAL) | <0.5 EU/device | Pass       |
| TEST                  | SPECIFICATIONS | CONCLUSION |
| Mouse Embryo Assay    | > 80%          | Pass       |
| MEA Control           | > 80%          | Pass       |



Karen Daemen  
Gynétics Medical Products N.V.

Signature date:  
3 May 2017