

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:

Pipette #3100

LOT number: 20170426**STERILIZATION date:** 29-04-2017**EXPIRY date:** 26-04-2021**STERILIZATION technique:** Gamma irradiation**STERILIZATION dose:** 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



K. Daemen
Gynetics Medical Products

SIGNATURE date:

17-05-2017