

Gynetics Medical Products CERTIFICATE OF QUALITY

ITEM name and code: # 4219 Emtrac Set 4.2

LOT NUMBER: 20170510

STERILIZATION DATE: 13-05-2017

EXPIRY DATE: 10-05-2021

STERILIZATION TECHNIQUE: Gamma Irradiation

STERILIZATION DOSE: 25 - 39 KGray

QUALITY CONTROL TESTING PROCEDURES

TEST	RESULT	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans
Gynetics Medical Products N.V.

31-05-17
Date