

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:

Embryo Transfer Cathether Emtrac A #4219 A

LOT number: 20170620**STERILIZATION date:** 24-06-2017**EXPIRY date:** 20-06-2021**STERILIZATION technique:** Gamma Irradiation**STERILIZATION dose:** 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans
Gynetics Medical Products

SIGNATURE date:

12-07-2017