

# Gynetics Medical Products

## CERTIFICATE OF QUALITY

**ITEM name and code:**  
Follicle Aspiration Needle #4551 FAS

**LOT number:** 20170704

**STERILIZATION date:** 08-07-2017

**EXPIRY date:** 04-07-2021

**STERILIZATION technique:** Gamma Irradiation

**STERILIZATION dose:** 25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

| TEST                  | SPECIFICATION  | CONCLUSION |
|-----------------------|----------------|------------|
| Endotoxin Assay (LAL) | <0.5 EU/device | Pass       |

| TEST               | SPECIFICATIONS | CONCLUSION |
|--------------------|----------------|------------|
| Mouse Embryo Assay | >= 80%         | Pass       |
| MEA Control        | >= 80%         | Pass       |



Greetje Hermans  
Gynetics Medical Products

**SIGNATURE date:**  
31-07-2017