

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:

Pipette #3125

LOT number: 20170705

STERILIZATION date: 08-07-2017

EXPIRY date: 05-07-2021

STERILIZATION technique: Gamma Irradiation

STERILIZATION dose: 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans
Gynetics Medical Products

SIGNATURE date:

31-07-2017