

# Gynetics Medical Products

## CERTIFICATE OF QUALITY

**ITEM name and code:**

Pipette #3100

**LOT number:** 20170712

**STERILIZATION date:** 15-07-2017

**EXPIRY date:** 12-07-2021

**STERILIZATION technique:** Gamma Irradiation

**STERILIZATION dose:** 25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans  
Gynetics Medical Products

**SIGNATURE date:**

16-08-2017