

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:

Intra-Uterine Insemination Cannula #4220 IUI

LOT number: 20170830**STERILIZATION date:** 02-09-2017**EXPIRY date:** 30-08-2021**STERILIZATION technique:** Gamma Irradiation**STERILIZATION dose:** 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans
Gynetics Medical Products

SIGNATURE date:

20-09-2017