

# Gynetics Medical Products

## CERTIFICATE OF QUALITY

**ITEM name and code:**

Embryo Transfer Catheter The Tulip #4000 The Tulip

**LOT number:** 20170906**STERILIZATION date:** 16-09-2017**EXPIRY date:** 06-09-2021**STERILIZATION technique:** Gamma Irradiation**STERILIZATION dose:** 25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans  
Gynetics Medical Products

**SIGNATURE date:**

04-10-2017