

# Gynetics Medical Products

## CERTIFICATE OF QUALITY

**ITEM name and code:**

Pipette #3125

**LOT number:** 20171004

**STERILIZATION date:** 07-10-2017

**EXPIRY date:** 04-10-2021

**STERILIZATION technique:** Gamma Irradiation

**STERILIZATION dose:** 25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans  
Gynetics Medical Products

**SIGNATURE date:**

18-10-2017