

# Gynetics Medical Products

## CERTIFICATE OF QUALITY

**ITEM name and code:**

Intra-Uterine Insemination Cannula #4220 IUI

**LOT number:** 20171004**STERILIZATION date:** 07-10-2017**EXPIRY date:** 04-10-2021**STERILIZATION technique:** Gamma Irradiation**STERILIZATION dose:** 25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans  
Gynetics Medical Products

**SIGNATURE date:**

18-10-2017