

CERTIFICATE OF QUALITY

ITEM name and code	#4000 S Memo TIS - Embryo Transfer Catheter
LOT NUMBER	20171009
EXPIRY DATE	9 October 2021
STERILIZATION DATE	14 October 2017
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass

Greetje Hermans
Gynetics Medical Products N.V.

Signature date:
13 December 2018