

Gynetics Medical Products CERTIFICATE OF QUALITY

ITEM name and code: Follicle Aspiration Needle #4551
LOT NUMBER: 20171022
STERILIZATION DATE: 11-11-2017
EXPIRY DATE: 2021-10-22
STERILIZATION TECHNIQUE: Gamma Irradiation
STERILIZATION DOSE: 25 - 39 KGray

QUALITY CONTROL TESTING PROCEDURES

TEST	RESULT	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans
Gynetics Medical Products N.V.



Date