

Gynetics Medical Products CERTIFICATE OF QUALITY

ITEM name and code: Follicle Aspiration Needle #4551

LOT NUMBER: 20171024

STERILIZATION DATE: 28-10-2017

EXPIRY DATE: 2021-10-24

STERILIZATION TECHNIQUE: Gamma Irradiation

STERILIZATION DOSE: 25 - 39 KGray

QUALITY CONTROL TESTING PROCEDURES

TEST	RESULT	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans
Gynetics Medical Products N.V.

01-12-2017
Date