

## CERTIFICATE OF QUALITY

<b>ITEM name and code</b>	#4000 L Memo - Embryo Transfer Catheter
<b>LOT NUMBER</b>	20180111
<b>EXPIRY DATE</b>	11 January 2022
<b>STERILIZATION DATE</b>	20 January 2018
<b>STERILIZATION TECHNIQUE</b>	Gamma Irradiation
<b>STERILIZATION DOSE</b>	25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

<b>TEST</b>	<b>SPECIFICATION</b>	<b>CONCLUSION</b>
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

  

<b>TEST</b>	<b>SPECIFICATIONS</b>	<b>CONCLUSION</b>
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans  
Gynetics Medical Products N.V.

Signature date:  
31 January 2018