

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:
Follicle Aspiration Needle #4551 FAS

LOT number: 20180123

STERILIZATION date: 27-01-2018

EXPIRY date: 23-01-2022

STERILIZATION technique: Gamma Irradiation

STERILIZATION dose: 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans
Gynetics Medical Products

SIGNATURE date:
21-02-2018