

CERTIFICATE OF QUALITY

ITEM name and code	#4225 Smooze - Intra-Uterine Insemination Catheter/Cannula Soft
LOT NUMBER	20180129
EXPIRY DATE	29 January 2022
STERILIZATION DATE	3 February 2018
STERILIZATION TECHNIQUE	Ethylene Oxide
RESULTS OF BI	No growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans
Gynetics Medical Products N.V.

Signature date:
21 February 2018