

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:
Handling Pipette #3125 Handling Pipette

LOT number: 20180207

STERILIZATION date: 10-02-2018

EXPIRY date: 07-02-2022

STERILIZATION technique: Gamma Irradiation

STERILIZATION dose: 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans
Gynetics Medical Products

SIGNATURE date:
28-02-2018