

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:
Follicle Aspiration Needle #4551 FAS

LOT number: 20180207

STERILIZATION date: 10-02-2018

EXPIRY date: 07-02-2022

STERILIZATION technique: Gamma Irradiation

STERILIZATION dose: 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

| TEST | SPECIFICATION | CONCLUSION |
|-----------------------|----------------|------------|
| Endotoxin Assay (LAL) | <0.5 EU/device | Pass |

| TEST | SPECIFICATIONS | CONCLUSION |
|--------------------|----------------|------------|
| Mouse Embryo Assay | >= 80% | Pass |
| MEA Control | >= 80% | Pass |



Greetje Hermans
Gynetics Medical Products

SIGNATURE date:
28-02-2018