

## Gynetics Medical Products CERTIFICATE OF QUALITY

**ITEM name and code:** #4164 Probet  
Endometrial Curette

**LOT NUMBER:** 20180214

**EXPIRY DATE:** 14 February 2022

**STERILIZATION DATE:** 17 February 2018

**STERILIZATION TECHNIQUE:** EO

**Results of sterility test with bio-indicators** No growth of organisms

Gynetics guarantees the following results due to a validated EO sterilization process.

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATIONS	Result
Ethylene oxide residue testing	<250 ppm	Pass
Ethylene chlorhydrin residue testing	<250 ppm	Pass
Suction function testing	<-110 mbars	Pass

  
\_\_\_\_\_  
Greetje Hermans  
Gynetics Medical Products N.V.

7-3-2018  
\_\_\_\_\_  
Date