

CERTIFICATE OF QUALITY

ITEM name and code	#4225 Smooze Medium - Intra-Uterine Insemination Catheter/Cannula Soft
LOT NUMBER	20180228
EXPIRY DATE	28 February 2022
STERILIZATION DATE	10 March 2018
STERILIZATION TECHNIQUE	Ethylene Oxide
RESULTS OF BI	No growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass

Greetje Hermans
Gynétics Medical Products N.V.

Signature date:
28 March 2018