

# Gynetics Medical Products

## CERTIFICATE OF QUALITY

**ITEM name and code:**

Intra-Uterine Insemination Cannula #4220 IUI

**LOT number:** 20180314**STERILIZATION date:** 17-03-2018**EXPIRY date:** 14-03-2022**STERILIZATION technique:** Gamma Irradiation**STERILIZATION dose:** 25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans  
Gynetics Medical Products

**SIGNATURE date:**

17-04-2018