

CERTIFICATE OF QUALITY

ITEM name and code	#4225 Smooze - Intra-Uterine Insemination Catheter/Cannula Soft
LOT NUMBER	20180314
EXPIRY DATE	14 March 2022
STERILIZATION DATE	16 April 2018
STERILIZATION TECHNIQUE	Ethylene Oxide
RESULTS OF BI	No growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans
Gynetics Medical Products N.V.

Signature date:
2 May 2018