

## CERTIFICATE OF QUALITY

<b>ITEM name and code</b>	#4225 Smooze Medium - Intra-Uterine Insemination Catheter/Cannula Soft
<b>LOT NUMBER</b>	20180418
<b>EXPIRY DATE</b>	18 April 2022
<b>STERILIZATION DATE</b>	21 April 2018
<b>STERILIZATION TECHNIQUE</b>	Ethylene Oxide
<b>RESULTS OF BI</b>	No growth of organisms

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass

Greetje Hermans  
Gynetics Medical Products N.V.

Signature date:  
9 May 2018