

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:

Soft Embryo Transfer Catheter #2000 Semtrac Set Short

LOT number: 20180418**STERILIZATION date:** 21-04-2018**EXPIRY date:** 18-04-2022**STERILIZATION technique:** ETO**STERILIZATION dose:** 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans
Gynetics Medical Products

SIGNATURE date:

09-05-2018