

Gynetics Medical Products CERTIFICATE OF QUALITY

ITEM name and code: Follicle Aspiration Needle #4551 FAS

LOT NUMBER: 20180424

STERILIZATION DATE: 28-04-2018

EXPIRY DATE: 24-4-2022

STERILIZATION TECHNIQUE: Gamma

STERILIZATION DOSE 25 - 39 KGray

QUALITY CONTROL TESTING PROCEDURES

TEST	RESULT	CONCLUSION
Endotoxin Assay (LAL)	<1.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans
Gynetics Medical Products N.V.

31-5-2018
Date