

## Gynetics Medical Products CERTIFICATE OF QUALITY

**ITEM name and code:** Follicle Aspiration Needle #4551 FAS

**LOT NUMBER:** 20180425

**STERILIZATION DATE:** 28-04-2018

**EXPIRY DATE:** 25-4-2022

**STERILIZATION TECHNIQUE:** Gamma

**STERILIZATION DOSE** 25 - 39 KGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	RESULT	CONCLUSION
Endotoxin Assay (LAL)	<1.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass

  
Greetje Hermans  
Gynetics Medical Products N.V.

31-5-2018  
Date