

## Gynetics Medical Products CERTIFICATE OF QUALITY

**ITEM name and code:** #4164 Probet  
Endometrial Curette

**LOT NUMBER:** 20180516

**EXPIRY DATE:** 16 May 2022

**STERILIZATION DATE:** 01 June 2018

**STERILIZATION TECHNIQUE:** EO

**Results of sterility test with bio-  
indicators** No growth of organisms

Gynetics guarantees the following results due to a validated EO sterilization process.

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATIONS	Result
Ethylene oxide residue testing	<250 ppm	Pass
Ethylene chlorhydrin residue testing	<250 ppm	Pass
Suction function testing	<-110 mbars	Pass



Greetje Hermans  
Gynetics Medical Products N.V.

25-6-2018  
Date