

CERTIFICATE OF QUALITY

ITEM name and code	#4000 S Memo - Embryo Transfer Catheter
LOT NUMBER	20180521
EXPIRY DATE	21 May 2022
STERILIZATION DATE	26 May 2018
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans
Gynétics Medical Products N.V.

Signature date:
30 July 2018