

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:

Embryo Transfer Catheter #4000 The Tulip

LOT number: 20180523**STERILIZATION date:** 26-05-2022**EXPIRY date:** 23-05-2022**STERILIZATION technique:** Gamma Irradiation**STERILIZATION dose:** 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans
Gynetics Medical Products

SIGNATURE date:

15-06-2018