

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:

Handling Pipette #3100 Handling Pipette

LOT number: 20180529**STERILIZATION date:** 08-06-2018**EXPIRY date:** 29-05-2022**STERILIZATION technique:** Gamma Irradiation**STERILIZATION dose:** 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass

Greetje Hermans
Gynetics Medical Products**SIGNATURE date:**
25-06-2018