

# Gynetics Medical Products

## CERTIFICATE OF QUALITY

**ITEM name and code:**

Follicle Aspiration Needle #4551 FAS

**LOT number:** 20180619**STERILIZATION date:** 22-06-2018**EXPIRY date:** 19-06-2022**STERILIZATION technique:** Gamma Irradiation**STERILIZATION dose:** 25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass

Greetje Hermans  
Gynetics Medical Products**SIGNATURE date:**

30-07-2018