

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:

Follicle Aspiration Needle #4551 FAS

LOT number: 20180704**STERILIZATION date:** 09-07-2018**EXPIRY date:** 04-07-2022**STERILIZATION technique:** Gamma Irradiation**STERILIZATION dose:** 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

| TEST | SPECIFICATION | CONCLUSION |
|-----------------------|----------------|------------|
| Endotoxin Assay (LAL) | <0.5 EU/device | Pass |

| TEST | SPECIFICATIONS | CONCLUSION |
|--------------------|----------------|------------|
| Mouse Embryo Assay | >= 80% | Pass |
| MEA Control | >= 80% | Pass |



Greetje Hermans
Gynetics Medical Products

SIGNATURE date:

08-08-2018