

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:

Soft Embryo Transfer Catheter #2000 Semtrac Set 4.2

LOT number: 20180711**STERILIZATION date:** 19-07-2018**EXPIRY date:** 11-07-2022**STERILIZATION technique:** ETO**STERILIZATION dose:** 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass



Greetje Hermans
Gynetics Medical Products

SIGNATURE date:

22-08-2018