

Gynetics Medical Products

CERTIFICATE OF QUALITY

ITEM name and code:

Follicle Aspiration Needle #4551 FAS

LOT number: 20180808**STERILIZATION date:** 09-08-2018**EXPIRY date:** 08-08-2022**STERILIZATION technique:** Gamma Irradiation**STERILIZATION dose:** 25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	>= 80%	Pass
MEA Control	>= 80%	Pass

Greetje Hermans
Gynetics Medical Products**SIGNATURE date:**

29-08-2018