

## CERTIFICATE OF QUALITY

<b>ITEM name and code</b>	#3125 Handling Pipette - Handling Pipette
<b>LOT NUMBER</b>	20180822
<b>EXPIRY DATE</b>	22 August 2022
<b>STERILIZATION DATE</b>	25 August 2018
<b>STERILIZATION TECHNIQUE</b>	Gamma Irradiation
<b>STERILIZATION DOSE</b>	25 - 39 kGray

### QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Greetje Hermans  
Gynetics Medical Products N.V.

Signature date:  
27 September 2018