


CERTIFICATE OF QUALITY

ITEM name and code	#4000 LS Memo - Embryo Transfer Catheter
LOT NUMBER	20180902
EXPIRY DATE	2 September 2022
STERILIZATION DATE	12 September 2018
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	25 - 39 kGray

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<0.5 EU/device	Pass

TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass



Evelien Sommen
Gynétics Medical Products N.V.

Signature date:
28 February 2019